

PEDIDO DE ESCLARECIMENTO - EMPRESA: AUTO SUTURE DO BRASIL LTDA - PE 21/2026 - MUNICIPIO DE CANTAGALO

1 mensagem

Carlos, Lara <lara.carlos@medtronic.com>

16 de setembro de 2026 às 12:59

Para: "licitacao@cantagalo.rj.gov.br" <licitacao@cantagalo.rj.gov.br>

Cc: "Marinho, Karina" <karina.b.marinho@medtronic.com>, "Santos, Zuleide" <zuleide.c.santos@medtronic.com>, "Galdino, Elisabete [BRA07]" <elisabete.galdino@medtronic.com>

Prezados, bom dia. A AUTO SUTURE DO BRASIL LTDA, inscrita no CNPJ/ME sob o nº 01.645.409/0003-90, apresenta o seguinte esclarecimento:

Analizando o EDITAL DE PREGÃO ELETRÔNICO SRP 985821-21/2026 – MUNICIPIO DE CANTAGALO com a intenção de apresentarmos nossa proposta de maneira correta, solicitamos a gentileza de nos esclarecer sobre os questionamentos abaixo:

os sensores Guardian Link **MMT-7020C1** apresentam validade de **12 meses** a partir da sua fabricação. Por se tratar de produto importado, é necessário levar em consideração que há um prazo para sua importação, logística e trâmites de desembaraço aduaneiro. Assim, o sensor, no seu momento de entrega, possuirá validade **pelo menos de 6 meses** da data de sua fabricação para comercialização no Brasil.

Pergunta: os sensores serão aceitos com prazo de validade de 180 (cento e oitenta) dias?

Obrigada e sendo assim, aguardamos retorno desta R. Administração.

Atenciosamente,

Lara de Lima Carlos

Estagiária de Licitação | Captação | Bid & Tenders

Medtronic

Tower Bridge Corporate

Av. Jornalista Roberto Marinho, 85 / 10º Floor | São Paulo, SP, 04576-010 | Brazil

lara.carlos@medtronic.com

medtronic.com | [Facebook](#) | [LinkedIn](#) | [Twitter](#) | [YouTube](#)

Medtronic

Engineering the extraordinary

[CONFIDENTIALITY AND PRIVACY NOTICE] Information transmitted by this email is proprietary to Medtronic and is intended for use only by the individual or entity to which it is addressed, and may contain information that is private, privileged, confidential or exempt from disclosure under applicable law. If you are not the intended recipient or it appears that this mail has been forwarded to you without proper authority, you are notified that any use or dissemination of this information in any manner is strictly prohibited. In such cases, please delete this mail from your

records. To view this notice in other languages you can either select the following link or manually copy and paste the link into the address bar of a web browser: <http://emaildisclaimer.medtronic.com>